CREATIVE BIOMART®
Assay Kit

Indican Test Kit

Product Information

Cat.No. Kit-2076

Product Overview

Indican test kit is a less hazardous modification of the original Obermeyer indican reagent. It provides an accurate, simple, fast (15 min), and inexpensive screening tool for the identification of protein putrefaction in the gastrointestinal tract.

Stability

12 months

Shipping

The kit is shipped and stored at room temperature. Protect from sunlight.

Size

20 tests

Kit Components

20 Vials of Reagent
20 Droppers

Target Species

Human

Detection method Obermeyer

Compatible Sample Types

Urine

Assay Protocol

Caution: Corrosive! The Indican Reagent contains hydrochloric acid. When performing the test, wear gloves, goggles and protective clothing.
In case of accidental contact with the reagent, immediately cover with baking soda and wash with

Tel: 1-631-559-9269 1-516-512-3133
Email:info@creative-biomart.org
Fax:1-631-938-8127
45-1 Ramsey Road, Shirley, NY 11967, USA

Indican Test Kit

ample amount of warm water.
1. Unscrew the Cap off the Indican vial.
2. Using the dropper, fill to the top "2-mL" mark and carefully add the 2 mL urine sample to the vial. Avoid spillage.
3. Replace cap, securely close the vial. Immediately invert the vial five times. Invert the vial again five times after 5, 10, and 15 minutes.
4. Compare the color in the top solvent layer with the Indican Color Chart below. Record results and dispose the materials properly.

Analysis

Expected Results
Clear or Light Yellow 0: Normal
Greenish Blue 1: Low Positive
Light Blue 2: Medium Positive
Blue 3: High Positive
Dark Blue 4: Very High Positive

Tel: 1-631-559-9269 1-516-512-3133
Email:info@creative-biomart.org
Fax:1-631-938-8127
45-1 Ramsey Road, Shirley, NY 11967, USA